

**E003:****Refrigerators and freezers**

<b>PQS code:</b>	<b>E003/061</b>
<b>Type of appliance:</b>	Vaccine/waterpacks freezer
<b>Manufacturer's reference:</b>	DW-25W300
<b>Manufactured in:</b>	China; People's Republic of
<b>Company:</b>	Qingdao Aucma Global Medical Co.,Ltd.
<b>Address:</b>	315 Qianwangang Road, Economic & Technological Development Zone, Qingdao, China
<b>Telephone:</b>	+86 532 86763602
<b>Email:</b>	<a href="mailto:lishouda@aucma.com.cn">lishouda@aucma.com.cn</a> <a href="mailto:info@aucma.com.cn">info@aucma.com.cn</a> <a href="mailto:chenzhipeng@aucma.com.cn">chenzhipeng@aucma.com.cn</a>
<b>Web address:</b>	<a href="http://www.aucma.com.cn">http://www.aucma.com.cn</a>

**Specifications**

<b>Climate zone:</b>	Hot	<b>Min rated ambient temp:</b>	+5°C
<b>Refrigerant:</b>	R134a	<b>Energy source:</b>	Electric Mains 220-240V 50/60Hz
<b>Appliance tested at:</b>	+43°C	<b>Ext dimensions (HxLxD)</b>	1226x790x945 mm
<b>Performance at:</b>	<b>+43°C</b>	<b>Fuel and cycle type:</b>	Electric - compression

<b>REFRIGERATOR</b>	<b>FREEZER</b>	<b>Freeze protection: N/A</b>
---------------------	----------------	-------------------------------

<b>Vaccine storage capacity: (Litres)</b>	240L		
<b>Gross volume (Litres):</b>		<b>Gross volume:</b>	300L
<b>Holdover time (hours):</b>	58hr 36Mins	<b>Waterpack freezing capacity:</b>	38.3kg/24h
<b>Energy consumption, stable running (kWh/24 hours):</b>	3.37kWh/24h		44.3kg
<b>Energy consumption, cool down test (kWh/24 hours):</b>	3.54kWh/24h	<b>Waterpack storage capacity:</b>	
		<b>Energy consumption during freezing:</b>	

**Comments:****Accessories:**

<b>Spare parts (ref, price):</b>	Included in the UNICEF Supply Division spare parts kit: 1 No. Compressor 1 No. Electronic thermostat 1 No. Fan
----------------------------------	---

Other parts are available as spares on request.

<b>Shipping volume:</b>	0.92 m3	<b>Shipping weight:</b>	75kg
<b>Price year</b>	2016	<b>Incoterms</b>	
<b>Price</b>	1-99 units, USD 540	100-199 units, USD 520	>200 units, USD 490
<b>Quality standard:</b>	- ISO 9001:2008 -	<b>Specification ref.</b>	
<b>Verification report:</b>	WB-14-041	<b>Verification laboratory:</b>	CHEARI
<b>Current PQS status:</b>	<i>pre-qualified: 02 Nov 2015</i>	<b>Validity until:</b>	<i>May, 2021</i>

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.